

附表三

檢定用法碼最大允許誤差表

單位:mg

待檢定衡器 自備法碼等級 標稱值	Ⅱ / F ₂	Ⅲ / M ₂	Ⅳ / M ₃
1000 kg	16000	160000	500000
500 kg	8000	80000	250000
200 kg	3000	30000	100000
100 kg	1600	16000	50000
50 kg	800	8000	25000
20 kg	300	3000	10000
10 kg	160	1600	5000
5 kg	80	800	2500
2 kg	30	300	1000
1 kg	16	160	500
500 g	8.0	80	250
200 g	3.0	30	100
100 g	1.6	16	50
50 g	1.0	10	30
20 g	0.8	8.0	25
10 g	0.6	6.0	20
5 g	0.5	5.0	16
2 g	0.4	4.0	12
1 g	0.3	3.0	10

註：表中允許誤差係指正負誤差